

**EU Declaration of Conformity**

Manufacturer according to Regulation 2017/745	Schülke & Mayr GmbH Robert-Koch-Str. 2 22851 Norderstedt Germany		
Registration Number acc. to Art. 31 2017/745	DE-MF-000005701		
Product Name	aspirmatic®		
Basic UDI-DI Code acc. to Art. 26 2017/745 Intended Purpose	4032651BSC00000018AD D99 disinfectant for suction units		
Risk Class according to Regulation 2017/745	II a Annex VIII rule 16		
Standards applied	EN ISO 13485 additional standards see technical documentation Schülke & Mayr GmbH		
Notified Body	DQS Medizinprodukte GmbH August-Schanz-Str. 21 60433 Frankfurt am Main Germany No.: 0297		
Conformity Assessment Procedure according to Regulation 2017/745	Annex IX	Chapter I, II section 4 and III	
Certificates	Annex IX EN ISO 13485	004567 MDR2017Q 004567 MDR2017B 004567 MP2016	
Version	1-0		

Schülke & Mayr GmbH herewith declares that the device covered by this declaration is in conformity with the Regulation 2017/745 concerning medical devices.

Schülke & Mayr GmbH declares that Schülke & Mayr GmbH bears the sole responsibility for issuing this declaration

Norderstedt 15.06.2023  
ppa.

  
Dr. Sven Pflüger  
Schülke & Mayr GmbH  
Chief Commercial Officer

15.06.2023  
ppa.

  
Lars Lemke  
Schülke & Mayr GmbH  
Chief Operating Officer