

Annex VII EC Declaration of Conformity

Manufacturer Name and Address: SpofaDental a.s.
IČ 63999447
Markova 238, 506 01 Jičín, Czech Republic

Technical File Name/Number: TD A118/3

Product Tradename: Kavitan Cem

Device Category: Glass-Ionomer Luting Cement

Device Identification: see Attachment 1

Classification and Rule: Class IIa, Rule 5 per Annex IX of the Council Directive 93/42/EEC

Notified Body: TÜV NORD CERT GmbH
Notified Body Number: 0044
Conformity Assessment Procedure & Certificate Issued: Annex II, CE certificate 44 232 141788, Edition 3

Declaration Statement:

We hereby declare that the above-mentioned device(s) comply with Council Directive 93/42/EEC.

Regulatory Affairs Signature:

10 -11- 2020

Issue date:

Name: Mgr. Jaroslav Adolf
Title: Quality Manager

valid until 2024-05-26

SpofaDental a.s.
Markova 238, CZ - 506 01 Jičín
IČO: 63999447
DIČ: CZ63999447
-19-

**TD A118/3 Kavitan Cem - Attachment 1 to Annex VII EC
Declaration of Conformity**

REF	Description
4113270PE	KAVITAN CEM, 20g PLV + 15g LIQ
4113273PE	KAVITAN CEM, 60g PLV + 45g LIQ (triple pack)
4113279PE	KAVITAN CEM, 15g LIQ