



# DECLARATION OF CONFORMITY

ACCORDING TO (EU) 2017/745 MEDICAL DEVICE REGULATION

## EU Representative

SUNGO Europe B.V.  
Olympisch Stadion 24, 1076DE  
Amsterdam, Netherlands  
SRN: NL-AR-000000247

## Conformity Assessment

### Conformity Assessment Procedure

Article 19, Annex II and Annex III according to REGULATION (EU) 2017/745.

### Applicable Standards

EN ISO14971:2019  
EN ISO15223-1:2016  
ISO10993-1:2018  
EN ISO10993-5:2009  
EN ISO10993-10:2010  
EN 1041:2008+A1:2013  
EN 62366-1:2015

## Remark

The declaration of conformity is valid in connection with the release technical document [CE/MDR-TRIBEST06](#).

All the supporting documentation is retained at the premises of the manufacturer.

The Declaration of Conformity is exclusively under the sole responsibility of the manufacturer.

## Manufacturer

Name: TRIBEST DENTAL PRODUCTS CO., LTD.  
Address: NO.98 ZHONGLING ROAD, SANMAO STREET, YANGZHONG, ZHENJIANG, JIANGSU, CHINA  
SRN: CN-MF-000027307

## Product Information

Name: Disposable Suction Tip  
Model: 11-10001 ~ 11-10040  
GMDN: 37434  
Basic UDI-DI: 69748153711-10034PU  
Classification: Class I

*On behalf of SUNGO Europe office, I confirmed we are EU REP of the company who issue this document.*



## Declaration

*Authorized Signature (S)*

We herewith declare that the above-mentioned products meet the requirements of Medical Device Regulation (EU) 2017/745 and the applicable standards above.

Signature:  Date: Mar 1<sup>st</sup>, 2021

*Zhou Jiahong*  
Position: General Manager Place: Yangzhong