



MICRO-NX Co., Ltd.
22, Maeyeo-ro 1-gil, Dong-gu, Daegu, 41059 Republic of Korea

May 26, 2023

Confirmation Letter Reference: CLNB1639 – KR/SEL/Y-PC/14366

To whom it may concern,

Confirmation of receipt of a formal application and conclusion of written agreement in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.

This letter confirms that, SGS Belgium NV, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number 1639 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer

MICRO-NX Co., Ltd.
22, Maeyeo-ro 1-gil, Dong-gu
41059 Daegu
Republic of Korea
SRN Number (if available): KR-MF-000026311

JaviTech e.K.
Sachsenhausener Str. 16
65824 Schwalbach a. Ts.
Germany

The devices covered by the formal application and the written agreement mentioned above are shown at the end of this letter.


In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired prior to the publication of the Regulation EU 2023/607 on 15th March 2023, this letter also confirms that:

- The manufacturer submitted the MDR application and signed the written agreement by the date of MDD/AIMDD certificate expiry;
- The certificates expired after 26th May 2021 by course of time and were valid at the date of their expiry neither having been suspended nor withdrawn.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3 of MDR (as amended by EU 2023/607), are shown below:

- 26th May 2026 for Class III custom-made implantable devices
- 31st December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31st December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31st December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body SGS Belgium NV 1639,



Ian How
PP

Virginie SILORET
Global Medical Device Certification Manager
Email: Virginie.siloret@sgs.com
Phone: +41 22 739 98 58

Devices covered by this letter:

Device name / Basic UDI-DI	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Electric Micromotor and Control Unit for Dental Handpiece (Model: NXOP-100E, NXHW-100E, EL-B40S, EL-B40L, EL-B40I, EL-B40M, EL-B40P)	Class IIa	N/A	NB1639 KR19/81826227

Device name / Basic UDI-DI	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Dental Torque Driver for Implant Surgery (Model: MEG-TORQ, OSM-TORQ, TD-C60, TD-C30);	Class IIa	N/A	NB1639 KR19/81826227
Dental Engine for Implant Surgery (Model: ISE-170, ISE-170L, ISE-170C, ISE-270M, ISE-270C);	Class IIa	N/A	NB1639 KR19/81826227

Confirmation Letter Revision History

Date	NB internal reference traceable to each version of the letter	Action
2023/05/26	Version 1	Initial issue

SGS NB1639 - Confirmation letter Regulation (EU) 2023/607