



Ultradent Products GmbH • Am Westhover Berg 30 • 51149 Cologne

Andrew Gould BDS
Tel +49 (0) 2203-3592-234
Fax +49 (0) 2203-3592-22
E-mail:
andrew.gould@ultradent.com

Cologne, 28.11.2022

Classification information

Indications for use:

LC Block-Out Resin is used to create a reservoir for tooth whitening gels during the production of individual trays and may be used in other procedures in the dental laboratory, model and die repairs for example.

Description:

LC Block-Out Resin is a light-cured resin which is neither used on or near patients but only for procedures in the dental laboratory. There is therefore no risk to the patient if this product does not function correctly.

Conclusion:

We hereby certify that LC Block-Out Resin, manufactured by Ultradent Products Inc., located at 505 West Ultradent Drive (10200 South) South Jordan, UT 84095, USA is a dental laboratory item only and is not used on or near patients and so does not fulfil the definition of a Medical Device according to EEC 92/431 (MDD), EU 2017/745(MDR).



Andrew Gould BDS, Dentist / Zahnarzt

Regulatory Affairs and Training Coordinator

Tel: +49 2203 – 35 92 234

Fax: +49 2203 – 35 92 22

Ultradent website | Opalescence website



Ultradent Products GmbH
Am Westhover Berg 30
51149 Köln
Tel.: 0 22 03/35 92 - 0
Fax: 0 22 03/35 92 - 22
E-mail: InfoDE@ultradent.com