



## **Declaration of Conformity**

Manufacturer Medit Corp.

F9, F10, 8, Yangpyeong-ro 25-gil, Yeongdeungpo-gu, Seoul, 07207, Rep. of Korea

Tel +82-2-2193-9600

Manufacturer SRN KR-MF-000007132

EC Representative Meditrial Srl

Via Po 9 00198, Rome Italy

Tel +39 06 45429780 ecrep@meditrial.eu

Description of Device

Brand(Trade) Name

Intraoral Scanner i700, i600

Model

MD-IS0200, MD-IS0100

Intended Use The i700 & i600 systems are a dental 3D scanner intended to be used to digitally

record topographical characteristics of teeth and surrounding tissues. The systems produce 3D scans for use in computer assisted design and manufacturing of dental

restorations.

Basic UDI-DI 88000267MD-IS0200KK, 88000267MD-IS0100KE

EC Regulation Medical Device Regulation (EU) 2017/745

GMDN code 63669

Classification Class I, according to the rules 13 in Annex VIII

Conformity assessment EC

route

EC conformity declaration according to Annex II and Annex III of the Regulation MDR

2017/745

## Standards applied

EN 60601-1:2005 (Third Edition) + CORR. 1:2006 + CORR. 2:2007 + A1:2012

EN 60601-1-2:2014

EN ISO 10993-5:2009, 10993-10:2013, 10993-11:2018

EN 62366:2008 EN 62304:2006 EN ISO 14971:2012

This declaration of conformity is issued under the sole responsibility of Medit Corp.

We hereby declare that he medical device(s) specified above meet the provision of the Regulation (EU) MDR 2017/745 for medical devices. This declaration is supported by the Quality System approval to ISO 13485 issued by BSI Group. All supporting documentation is retained at the premises of the manufacturer.

Date of issue: MAR 03, 2022

Signature

Name : GYU BUM, KO

Title : CEO

AUTHORIZED SIGNATURE

MEDIT CORP.