

Declaration of Conformity

We hereby declare that the medical device, due to its design and the version we have placed on the market, complies with the basic requirements of Regulation (EU) 2017/745 considering article 120 (3). The manufacturer is solely responsible for issuing this declaration of conformity.

Manufacturer	METASYS Medizintechnik GmbH · Florianistrasse 3, 6063 Rum bei Innsbruck, Austria
SRN	AT-MF-000014438
Basic UDI-DI	++D98410022KD
Product trade name	GREEN&CLEAN SK
Article number / product designation	see Annex I
Intended purpose	Solution for disinfection of surfaces of non-invasive medical devices.
Classification	Risk class IIa (according to regulation (EU) 2017/745 Annex VIII)
Applied common specifications and other Union standards / legislations	see Annex 2
Notified body	TÜV NORD CERT GmbH (CE0044)
Certificate number	44 911 220982 valid until 08/08/2029
Conformity assessment	Regulation (EU) 2017/745, Article 19 and Annex II, III

Place, date
Rum, 08.08.2024


Bruno Pregonzer
Managing Director

Valid from: 08.08.2024

Valid until: Until the change of one of the products specified in Annex I, but at the latest until: 08.08.2029

Annex I - List of article numbers

UDI-DI - article number - product trade name

+D9841120000340 - 112000034 - GREEN&CLEAN SK 750 ml

+D9841120000350 - 112000035 - GREEN&CLEAN SK 5 l

+D9841120000470 - 112000047 - GREEN&CLEAN SK ATMOS 750 ml

n.a. / retail kit - 122000050 - GREEN&CLEAN SK 750 ml, 6 bottles, 2 foamers

+D9841120000350 - 122000051 - GREEN&CLEAN SK 5 l, 1 canister

n.a. / retail kit - 122000064 - GREEN&CLEAN SK 750 ml, 1 bottle, foamer

n.a. / retail kit - 122000067 - GREEN&CLEAN SK ATMOS 750 ml, 1 bottle, foamer

Annex II - Applied common specifications and other Union standards / legislations

This present declaration is also in conformity with the following Union standards/legislations and Common Specifications (CS):

Applied common specifications

not applicable

Standard/Legislation - Description

EN 13727 - Chemical disinfectants and antiseptics - Quantitative suspension test for the evaluation of bactericidal activity in the medical area
- Test method and requirements

EN 13624 - Chemical disinfectants and antiseptics - Quantitative suspension test for the determination of fungicidal or levurocidal activity in the field of human medicine - Test method and requirements

EN 16615 - Chemical disinfectants and antiseptics - Quantitative test method for the evaluation of bactericidal and yeasticidal activity on non-porous surfaces with mechanical action employing wipes in the medical area (4-field test) - Test method and requirements

EN 14476 - Chemical disinfectants and antiseptics - Quantitative suspension test for the evaluation of virucidal activity in the medical area - Test method and requirements

EN 16777 - Chemical disinfectants and antiseptics - Quantitative test on non-porous surfaces without mechanical action to determine virucidal activity in human medicine - Test method and requirements